

# WUJIANG PAILIXIN METAL PRODUCT CO; LTD

## Sterilization Certificate

General Information			
Sterilization No.		180208-1	
Customer name		PAINFUL PLEASURE	
Date of reception		2018.02.08	
Temperature of Environment		50°C	
Product name	Product type	Product Lot#	Cases
PRECISION CARTRIDGES	1203RLXT,1205RL,1207RS,1217RM,etc	0201	19CTNS
Sterilization process data			
Preconditioning time		1.5-2.5hrs	
Date of sterilization		2018.02.08	
Station number		3	
Batch number		180208	
Aeration time		7 Days	
Results of Physical process control			
The physical Acceptance criteria of the preconditioning the cycl and the aeration were in compliance with the validated specifications of cycle no.: <u>GBS</u>			
Microbiological process data			
The microbiological process control with biological indicators has been performed according to the BI sterility test sop(032005)			
No of biological indicators indicators 19 per chamber load.		BeiJin Four Ring Medical,Lot No.20170203	
Date of incubation		2018.02.08	
Incubation conditions		7 Days 37°C	
Result of microbiological process control			
[   ]No growth from Bacillus subtulis has been detected			
Remarks			
QA Approval			
Date of release		Signature of QA	
2018.02.09			

