

# WUJIANG PAILIXIN METAL PRODUCT CO; LTD

## Sterilization Certificate

General Information			
Sterilization No.	180718-1		
Customer name	PAINFUL PLEASURE		
Date of reception	2018.07.18		
Temperature of Environment	50°C		
Product name	Product type	Product Lot#	Cases
PRECISION CARTRIDGES	1207RL,1209M1,1007RL,1007RL,1009M1C,etc	0701	41CTNS
PLASTIC TUBE	30MM		10CTNS
Sterilization process data			
Preconditioning time	1.5-2.5hrs		
Date of sterilization	2018.07.18		
Station number	7		
Batch number	180718		
Aeration time	7 Days		
Results of Physical process control			
The physical Acceptance criteria of the preconditioning the cycl and the aeration were in compliance with the validated specifications of cycle no.: <u>GBS</u>			
Microbiological process data			
The microbiological process control with biological indicators has been performed according to the BI sterility test sop(032005)			
No of biological indicators indicators 19 per chamber load.	BeiJin Four Ring Medical, Lot No.20170203		
Date of incubation	2018.07.18		
Incubation conditions	7 Days 37°C		
Result of microbiological process control			
[    ] No growth from Bacillus subtilis has been detected			
Remarks			
QA Approval			
Date of release	Signature of QA		
2018.07.18			

