

WUJIANG PAILIXIN METAL PRODUCT CO; LTD

Sterilization Certificate

General Information			
Sterilization No.		190930-1	
Customer name		PAINFUL PLEASURE	
Date of reception		2019.09.30	
Temperature of Environment		50°C	
Product name	Product type	Product Lot#	Cases
STAINLESS STEEL PIN WITH TUBE	C1005RL 、 C1213MG 、 C1013RM、 1RL 0.20MM*B、 5RS 0.25MM etc	0901	142CTNS
PLASTIC RUBBER SLEEVE			2CTNS
Sterilization process data			
Preconditioning time		1.5-2.5hrs	
Date of sterilization		2019.09.30	
Station number		7	
Batch number		190901	
Aeration time		7 Days	
Results of Physical process control			
The physical Acceptance criteria of the preconditioning the cycl and the aeration were in compliance with the validated specifications of cycle no.: <u>GBS</u>			
Microbiological process data			
The microbiological process control with biological indicators has been performed according to the BI sterility test sop(032005)			
No of biological indicators indicators 19 per chamber load.		BeiJin Four Ring Medical, Lot No.20170203	
Date of incubation		2019.09.30	
Incubation conditions		7 Days 37°C	
Result of microbiological process control			
[] No growth from Bacillus subtilis has been detected			
Remarks			
QA Approval			
Date of release		Signature of QA	
2019.09.30			