

WUJIANG PAILIXIN METAL PRODUCT CO; LTD

Sterilization Certificate

General Information

Sterilization No.	181201-1		
Customer name	PAINFUL PLEASURE		
Date of reception	2018.12.01		
Temperature of Environment	50°C		
Product name	Product type	Product Lot#	Cases
STAINLESS STEEL PIN WITH TUBE	C1003RL、C1005RL、C1007、RL、C1009RL、C1207RL etc	1201	24CTNS

Sterilization process data

Preconditioning time	1.5-2.5hrs
Date of sterilization	2018.12.01
Station number	7
Batch number	181201
Aeration time	7 Days

Results of Physical process control

The physical Acceptance criteria of the preconditioning the cycel and the aeration were in compliance with the validated specifications of cycle no.:GBS

Microbiological process data

The microbiological process control with biological indicators has been performed according to the BI sterility test sop(032005)

No of biological indicators indicators 19 per chamber load.	BeiJin Four Ring Medical, Lot No.20170203
Date of incubation	2018.12.01
Incubation conditions	7 Days 37°C

Result of microbiological process control

[] No growth from Bacillus subtilis has been detected

Remarks

--	--

QA Approval

Date of release	Signature of QA
2018.12.01	

